



Short Tandem Repeat

Requestor: Aine Russell, Boston University

Samples Received: 09Sep20

STR Amplification Date: 16Sep20

Form SOP-89.01

Version 1.0

Sample Name	BU3 NGPT
Label on tube	82830
FGA	20, 24
TPOX	11, 11
D8S1179	14, 14
vWA	16, 18
Amelogenin	X, Y
Penta_D	12, 12
CSF1PO	10, 11
D16S539	9, 13
D7S820	9, 11
D13S317	12, 12
D5S818	10, 11
Penta_E	12, 14
D18S51	15, 16
D21S11	29, 32
TH01	7, 9.3
D3S1358	14, 15
Allelic Polymorphisms	26
Matches*	77672, 77674, 77673
Comments	

**Note: The STR profile of the following sample is an exact match for the given sample/samples.*



Short Tandem Repeat

Requestor: Aine Russell, Boston University
Samples Received: 09Sep20
STR Amplification Date: 16Sep20

ResultsThe genotypic profiles comprise a range of __26__ allelic polymorphisms across the 15 STR loci analyzed.

Interpretation: The concentration of DNA required to achieve an acceptable STR genotype (signal/ noise) was equivalent to that required for the standard procedure (~1 ng/amplification reaction) from human genomic DNA. These results suggests that the cells submitted correspond to the cell lines as named and were not contaminated with any other human cells or a significant amount of mouse feeder layer cells.

Sensitivity: Sensitivity limits for detection of STR polymorphisms unique to either this or other human cell lines is ~2-5%.

X

Tech #1
Characterization

X

Tech #2
Characterization

X

QA Review
Quality Assurance

Unless otherwise mutually agreed in writing, the services provided to you hereunder by WiCell Research Institute, Inc. ("WiCell") are governed solely by WiCell's Terms and Conditions of Service, found at www.wicell.org/privacyandterms. Any terms you may attach to a purchase order or other document that are inconsistent, add to, or conflict with WiCell's Terms and Conditions of Service are null and void and of no legal force or effect.

Raw data is available upon request.